

Follow-up scheme for patients with implanted neurostimulators.

Follow-up at:

- 2-3 days (in the hospital): control of wound and general status, program up date if necessary, explaining external programmer (EZ)
- 10 days: control of wound and general status, program up date if necessary
- 3 months: control use of device, battery status
- 6 months: same as above
- 1 year: same as above

After this point we see the patients once per year or on demand in case of loss of therapy effectiveness.